

Declaration for Bioreader® 7000

The Bioreader® 7000 devices are used to apply image processing technology to evaluate assays in Multiwell, Microtiter and Microfilter plates in the field of research as well as routine and diagnostic work.

They are general laboratory instruments and they do not have to be registered as IVD medical devices because they only count objects once the assay has been completed and they are not part of the actual biochemistry of the assays.¹



Picture 2: Bioreader(R) 7000 models



Picture 1: Typical Microfilter plate for the Elispot assay

This declaration refers to all **Bioreader® 7000 models**, especially in connection with **T-Spot® assays**² from Oxford Immunotec³:

- In this application the Bioreader® presents well images of a 96 well Microfilter plates with a PC plus monitor and submits count result. These results represent reactive T-cells by use of the Elispot assay. Therefore it is used for assisting the evaluation.
- The Bioreader® 7000 was verified with T-Spot®.TB reference plates from a reference laboratory of Oxford Immunotec⁴. Nevertheless, it is necessary to verify the Bioreader® on-site by use of the samples of the laboratory as differences in sample preparation and automation lead to different characteristics of the spots.

The counts need to be evaluated, corrected (if needed) and finally approved by a qualified medical technician.

- The usage of the Bioreader® device is only permitted by laboratory personnel who were trained by a medical product consultant and authorized by BIOSYS.
- Bioreader® 7000 is a semi-automatic device. The operator makes the final decision of the diagnosis.

¹ However, all Bioreader® 7000 models meet the applicable requirements of EU 2017/746 IVD-R ‚In-Vitro Diagnostics‘ as well as further standards / requirements / guidelines described on next page.

² Published declaration of conformity from Oxford Immunotec Ltd.: TB-DCEC-UK-0001-v20, T-SPOT.COVID-DC_EC_-UK-0001-v3 as well as DC-CMV-UKEC-V4.

³ Company Oxford Immunotec Ltd., 143 Park Drive, Milton Park, Abingdon, Oxfordshire, OX14 4SE, United Kingdom (authorized representative in EU: Oxford Immunotec (Ireland) Ltd., Unit 3d, North Point House, North Point Business Park, New Mallow Road, Cork, Ireland).

⁴ The related clinical performance study was done with the Bioreader® 7000 -E alpha (SN: 356).

| | |
|---|---|
| DF-13-E-3 | Title: Bioreader® 7000 Declaration |
| Changed : 27 th April 2023 Anthony Fry | Checked/Released: 27 th April 2023 Werner Freber |

BIOSYS Scientific Devices GmbH hereby declares that the Bioreader® 7000 meets the following **EU and ISO product standards (incl. standards for IVD product class A)**:

| Standards | Date | Description |
|---------------------|---------|--|
| ISO 12100 | 2010 | Safety of Machines (For the User) |
| DIN EN 61010-1 | 2020-03 | Safety Regulations for electrical measuring-, control- and laboratory Equipment – Part 1: General Requirement |
| DIN EN 61010-2-101 | 2017-10 | Safety Regulations for electrical measuring, control and laboratory Equipment – Part 2-101; Special Requirement for in vitro diagnostics (IVD) medical Devices |
| DIN EN 61326-1 | 2018 | Electrical measurement, control and laboratory Equipment EMC Requirement-Medical in vitro Part 1: General Requirement |
| DIN EN 61326-2-6 | 2018 | Electrical measurement, control and laboratory Equipment EMC Requirement. Part 2-6: Particular Requirement - -In vitro diagnostic medical Equipment (IVD) |
| IEC 61000-3-2 To -4 | | Electrical interference Resistance |

as well as following **EU and ISO system standards**:

| | | |
|---|------|--|
| DIN EN ISO 18113-3 :2021 EN ISO 15223-1:2020 | 2021 | In vitro diagnostic medical Devices – Provision of Information by the manufacturer – Part 3: Apparatus for in vitro diagnostic examination for use by qualified Staff. (EN ISO 15223-1:2020 Symbols) |
| EN ISO 14971:2019 | 2019 | Medical Devices – Application of Risk Management to Medical Devices Relevant publications: Appendix 2 |
| ISO 13485:2016/2021 | 2021 | Medical Devices – Quality Management Systems –Requirements for regulatory purposes |
| ISO 20916:2019 | 2019 | In vitro diagnostic medical devices – Clinical performance studies on human test material – Good practice (See Annex 2 publications [TD-306-D]) |
| EN 62304:2006 + A1:2015 | 2015 | Medical Device Software - Software Lifecycle Processes. Based partly on GAMP 4 (SDS and Lifecycle). |
| EN 62366-1:2015 | 2015 | Medical Devices - Part 1: Application of suitability for use of medical Devices. Applied and improved as far as relevant. |
| GAMP 4 | 2007 | The documentation according to the GAMP 4 specifications has been practiced since 2007. |

The person responsible for compliance with regulatory requirements: Werner Freber (Dipl.-Ing.), Kiefernweg 10, 61184 Karben, Germany.

Further technical documentation sees: TD-306-D deliverable information from manufacturer.

| | |
|---|---|
| DF-13-E-3 | Title: Bioreader® 7000 Declaration |
| Changed : 27 th April 2023 Anthony Fry | Checked/Released: 27 th April 2023 Werner Freber |